



America

CERTIFICATE

No. Z1 13 04 14544 087

Holder of Certificate: Fluke Corporation6920 Seaway Boulevard
Everett, WA 98203-9090
USA**Production Facility(ies):** 56588**Certification Mark:****Product:** Multimeters
(Handheld Digital Multimeter)**Model(s):** 83V, 87V, 88V

Parameters:	Rated Input Voltage:	9VDC (1X9V IEC 6LR61 battery)
	Measurement Category:	CAT III 1000V, CAT IV 600V
	Measuring Current:	10A max
	Measuring Voltage:	1000V max
	Protection Class:	II
	IP Protection:	IP30
	Temperature Ambient:	-20°C to +55°C

Tested according to: EN 61010-1:2010
EN 61010-2-030:2010

The product was tested on a voluntary basis and complies with the essential requirements. The certification mark shown above can be affixed on the product. It is not permitted to alter the certification mark in any way. In addition the certification holder must not transfer the certificate to third parties. See also notes overleaf.

Test report no.: PI1300226-000**Date,** 2013-04-16

Page 1 of 1

